ENO CALIBRATION GAS

Calibration of the NOA for gas-phase measurements is easily performed using this lightweight, disposable cylinder of 45 ppm NO in N2.
Part Number: CNDK 01400

POLYGEN Sp. z o.o. | ul. Portowa 16L/130 | 44-102 Gliwice
Tel.: +48 32 238 81 95 | Biuro: +48 725 725 373 | E-mail: polygen@polygen.com.pl