

PARTICLE FILTER - WHITE



For Gas Sampling Line. Should replace if pressure drop across reaction cell is greater than 0.1 Torr when filter is on versus when it is off, or should be changed every 6 months, whichever is sooner.

Part Number: CHFL 01410 A